

Data Acceptance Criteria - "Red Flag" Monitoring Volunteer Teams

	Temperature	pH	Dissolved Oxygen	Conductivity	Total Hardness
Data Acceptance Criteria					
Precision - Acceptance of Duplicates	± 1°C	± 0.5 pH Units	± 20%	± 10%	± 20%
Accuracy - Acceptance of Standards	N/A	± 0.5 pH Units	N/A	± 1%	± 20%
Splits - Comparison with certified lab	N/A	± 0.5 pH Units	N/A	± 20%	± 20%

Data that does not meet this criteria will NOT be entered in the CSI database!!